



Ref. :

الترقيم :

Date:

التاريخ :

دعوة دخول عطاء فرعي / لوازيم

رقم العطاء: (ح ٥٠٧)
الموضوع: توريد وتثبيت جهاز
Freeze Dryer

حضرة المتعهد المحترم،

تحية طيبة وبعد،

١- تحتاج جامعة الحسين بن طلال إلى اللوازم والخدمات المبينة تفصيلها في الجدول/الكشوف المرفقة،
٢- يرجى تقديم عرضكم للتوريد/أو التنفيذ وفق التعليمات والشروط العامة والخاصة والمواصفات المرفقة
بهذه الدعوة.

٣- يشترط للنظر في عرضكم أن يقدم على النموذج المرفق إن وجد، أو نموذج تقدمونه يشتمل على
البيانات المتعلقة باللوازم والخدمات المطلوبة، إذا توافر أكثر من سعر للمادة الواحدة، أو أدرجت
ملحوظات خاصة بالعرض.

- ٤- آخر موعد لقبول العرض قبل الساعة ٣ من يوم الموافق ٨/٢٧/٢٠١٧ م
٥- موعد فتح صندوق العطاءات الساعة ٣ من يوم الموافق ٨/٢٧/٢٠١٧ م
٦- لا ينظر في أي عرض يقدم بعد الموعد المحدد.
٧- مدة التسليم () يوماً تقويمياً من تاريخ تبليغ الإحالة.

وتفضلوا فائق الاحترام



جامعة الحسين بن طلال

رئيس لجنة العطاءات الفرعية / لوازيم

كارولين هادي

المواصفات المطلوبة لجهاز Freeze Dryer (lab scale)

- Condenser Temperature: -65°C or better
- Condenser volume: 4 litres minimum
- Efficient vacuum system (with corrosion resistant parts) that can reduce pressure to 0.01 mbar (or better)
- Heavy duty with high safety standards during operation
- Temperature and pressure display
- Ability to work with acidic or basic solutions is a privilege
- USA, European, Japanese or Korean origin is preferred.



